

flow is acceptable.

1.3 Data Quality Objectives 0.1 (e) -Adherence 0.1 (e) -to 0.1 (e) -the 0.1 (e) -require

6.1 Type S Pitot Tube.

intervals is part of a routine procedure, then comparative p measurements

6.7.1.2 A m733 re nm733.2 i(m733 reu) -0.((m733 re() -0.2 o) -0.1 f) -0.2 () -0.

flow system should have the capacity to generate at least four distinct, time-invariant test-section velocities covering the velocity range from 180 to 1,500 m/min (600 to 5,000 ft/min), and calibration data shall be taken at regular velocity intervals over this range (see References 9 and 14 in Section 17.0 for details).


that the entry port surrounding the tube is properly sealed.

10.1.3.4 Read p_{std} , and record its value in a data table similar to the one shown in Figure 2-9. Remove the **standard pitot tube** from the duct, and disconnect it from the manometer. Seal the standard entry port.

10.1.3.5 Connect the standard pitot tube to the manometer.

in.) are not ordinarily used for isokinetic sampling at velocities around 910 m/min (3,000 ft/min), which is the calibration velocity. Note also that it is not necessary to draw an isokinetic sample during calibrations. () -0.2 (s) -0.1 (e) -0.1 (e)

thermometer, or equivalent, as a reference. Alternatively, either a thermocouple and a potentiometer (calibrated against thermometric fixed points) or a bath and a boiling barometric pressure) may be used. For tem



Eq. 2-1

12.3 Calculate D_e as follows:

Eq. 2-4

12.4.4 Calculate F , the average deviation. $r = 0.3$, $r = 0.3$, $\sigma_C = 0.1$, $r = 0.1$, $t = 0$.

Inc. 1947.

6. Fluid meters - Their Theory and Ap



Figure 2-4. Proper temperature sensor placement to prevent interference; D

